

CBE Safety Inspection Report

Date:
Room:
PI:

S* I N/A

Housekeeping

- General housekeeping
- Clear evacuation routes (28 in)
- No tripping hazards
- No food/eating/drinking in lab areas

Signage

- Call-in sticker on door up to date
- Laser sticker on door
- Compressed gas sticker on door
- Responder hazard warnings on door
- Signs for fire extinguisher locations
- Other warning signs

Personnel

- Wearing safety glasses
- Appropriate dress
- CHP Training
- Chemical Hazardous Waste Training
- Right to Know Training

Safety Supplies

- Spill Kit
- First Aid Kit
- Visitor Safety Glasses

Records and Information

- CHP available
- MSDSs available
- SOPs available
- Waste accumulation area inspections
- Waste management guidelines
- Medical waste disposal records

S I N/A

Chemicals

- All containers legibly labeled
- All containers capped
- Chemicals appropriately segregated
- Gas cylinders chained
- Unused gas cylinders capped

Safety Equipment

- Eye-wash station functional
- Eye-wash station access
- Hoods inspected
- Hood sashes at proper height
- Hoods free of clutter
- Proper sharps disposal
- Proper glass disposal containers

Fire and Electrical Safety

- Fire extinguishers inspected
- Proper ceiling clearance (18/24 in)
- Proper electrical panel clearance
- No ceiling tiles missing
- No exposed electrical connections
- No extension cords
- No daisy-chained power strips
- GFI on electric circuits near water
- Amount of flammables within limits

Hazardous Waste

- Waste containers properly labeled (including start date)
- Wastes appropriately segregated
- Waste containers capped
- Designated waste accumulation area
- Regulated medical waste containers

Other comments:

* S = Satisfactory. I = Could be Improved, N/A = Not Applicable